

Name: Wuxi Institute of Inspection, Testing and Certification (Wuxi Institute of Metrology and Testing/Wuxi Center of Fiber Inspection)

Address: No. 200, Linghu Avenue, Xinwu District, Wuxi, Jiangsu, China

Registration No. CNAS L0260

CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT
Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS
SCHEDULE OF ACCREDITATION CERTIFICATE

Effective Date: 2023-06-19 Expiry Date: 2028-09-16

SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
I 、 Mechanical measuring instrument							
1	Vibration Displacement Transducer	Displacement	Verification Regulation of Vibration Displacement Transducer JJG 644	Static:(0~25)mm	$U_{\text{rel}}=0.3\%$		
				Dynamic:(0.01~5)mm, (0.4~500)Hz	$U_{\text{rel}}=1.5\%$		
2	Linear Accelerometer	Linear acceleration	Calibration Specification for MEMS Linear Accelerometers JJF1427	($>9.8\sim450$)m/s ² , (32~330)r/min	$U_{\text{rel}}=5 \times 10^{-4}$	中国合格评定国家认可委员会 认可专用章	
				(-9.8~-1.7 $\times 10^{-4}$)m/s ² , (-55~85)°C	$U_{\text{rel}}=2 \times 10^{-5}$		
				(1.7 $\times 10^{-4}$ ~9.8)m/s ² , (-55~85)°C	$U_{\text{rel}}=2 \times 10^{-5}$		
3	Vibratory Transmitter	Acceleration	Calibration Specification for Vibratory Transmitter JJF(Su)235	(0.1~100)m/s ² , (0.4~4000)Hz	$U_{\text{rel}}=1.5\%$		

No. CNAS L0260

第 1 页 共 5 页

The scope of the accreditation in Chinese remains the definitive version.



Nº	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Velocity		(0.1~100)mm/s, (0.4~1000)Hz	$U_{\text{rel}}=1.5\%$		
		Displacement		(0.01~5)mm, (0.4~500)Hz	$U_{\text{rel}}=1.5\%$		
4	Electromagnetic Sensor of Rotational Speed	Rotational Speed	C. S. for Electromagnetic Sensor of Rotational Speed JJF1871	(10~100)r/min	$U=0.03\text{r}/\text{min}$		
				(>100~8000)r/min	$U_{\text{rel}}=0.01\%$		
5	Speed and Mileage Meter for Standard Equipment	Rotate Speed	V.R.of Speed and Mileage Meter for Standard Equipment JJG779	(10~4000)r/min	$U_{\text{rel}}=0.6 \times 10^{-4}$		
6	Tachometer	Rotate Speed	V.R.of Tachometers JJG105	(10~40000)r/min	$U_{\text{rel}}=0.02\%$		
7	Aneroid Barograph and Aneroid Barometer	Pressure	V.R.of Aneroid Barograph and Aneroid Barometer JJG272	(500~1050)hPa	$U=0.4\text{hPa}$		
8	Equipment Standard for Revolution Speed	Rotate Speed	V.R.of Standard Equipment for Revolution Speed JJG326	(10~40000)r/min	$U_{\text{rel}}=0.9 \times 10^{-4}(k=3)$		
9	Digital Barometers	Pressure	V.R.of Digital Barometers JJG1084	(10~1200)hPa	$U=(0.2\sim0.4)\text{hPa}$		
10	Accelerometer	Acceleration	V.R.of Piezoelectric Accelerometer JJG233	(0.1~100)m/s ² , (0.4~20)Hz	$U_{\text{rel}}=1.4\%$		
				(0.1~1000)m/s ² , (>20~4000)Hz	$U_{\text{rel}}=1.5\%$		
11	Measuring Vibration Instruments	Acceleration	V.R.of Vibration Meters JJG676	(0.1~100)m/s ² , (5~2000)Hz	$U_{\text{rel}}=2.3\%$		
		Velocity		(0.1~100)mm/s, (5~1000)Hz	$U_{\text{rel}}=2.3\%$		
		Displacement		(0.01~5)mm, (5~500)Hz	$U_{\text{rel}}=2.3\%$		

No. CNAS L0260

第 2 页 共 5 页

The scope of the accreditation in Chinese remains the definitive version.



Nº	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
12	Portable Vibration Calibrator	Acceleration	V.R.of Portable Vibration Calibrator JJG1062	(0.1~1000)m/s ²	$U_{\text{rel}}=1.2\%$		
		Frequency		(5~4000)Hz	$U_{\text{rel}}=0.06\%$		
13	Photoelectric Belt Tension Meters	Frequency	Calibration Specification for Photoelectric Belt Tension Meters JJF(S) 204	(10~50)Hz	$U=0.1\text{Hz}$		
				(>50~600) Hz	$U=0.2\text{Hz}$		
II、Acoustic measuring instrument							
1	Personal Sound Exposure Meter	sound exposure	Personal Sound Exposure Meters JJG 980	Absolute acoustic sensitivity:10Pa ² h	$U=0.9\text{Pa}^2\text{h}$		
				Weighted, Steady State Response:(0.1~99.99)Pa ² h, (63~8000)Hz	$U_{\text{rel}}=6\%$		
				Short duration signal response:(0.1~99.99)Pa ² h, (63~8000)Hz	$U=0.12\text{Pa}^2\text{h}$		
2	Microphone Preamplifier	Frequency Response	Calibration Specification for Microphone Preamplifiers JJF1137	(-10~10)dB, (10 Hz~20 kHz)	$U=0.5\text{dB}$		
3	Dynamical Signal Analyzer	Frequency	V.R. of Dynamical Signal Analyzer JJG834	0.1Hz~200kHz	$U_{\text{rel}}=6 \times 10^{-5}$	中国合格评定国家认可委员会 认可证书专用章	
		Voltage		1mV~10V	$U_{\text{rel}}=0.6\%$		
4	*Pure tone audiometer	Hearing level zero	V.R.of Pure tone audiometer JJG388	Air conduction : (-10~120) dB, 125Hz~8kHz	$U=0.9\text{dB}$	中国合格评定国家认可委员会 认可证书专用章	
				Bone conduction : (-10~70) dB, 125Hz~8kHz	$U=1.7\text{dB}$		



No. CNAS L0260

第 3 页 共 5 页

The scope of the accreditation in Chinese remains the definitive version.

Nº	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
5	Sound level meter	Sound Pressure Level	V.R.of Sound Level Meters JJG188	Sound Signal:(30~130)dB, (10Hz~4kHz)	$U=0.6\text{dB}$		
				Sound Signal:(30~130)dB, (>4~10)kHz)	$U=0.7\text{dB}$		
				Sound Signal:(30~130)dB, (>10~20)kHz	$U=1.0\text{dB}$		
				Sine Signal:(30~130)dB, (31.5Hz~12.5kHz)	$U=0.2\text{dB}$		
				Toneburst Signal:(30~130)dB, (0.25~1000)ms	$U=0.3\text{dB}$		
				F:(20.0~50.0)dB/s	$U=3.2\text{dB/s}$		
				S:(1.0~8.0)dB/s	$U=0.3\text{dB/s}$		
6	Noise Level Statistical Analyzers	Sound Pressure Level	V.R.of Noise Level Statistical Analyzers JJG778	Sound Signal:(30~130)dB, (10Hz~4kHz)	$U=0.6\text{dB}$		
				Sound Signal:(30~130)dB,(>4kHz~10kHz)	$U=0.7\text{dB}$		
				Sound Signal: (0~130)dB, (>10~20)kHz	$U=1.0\text{dB}$		
				Sine Signal: (30~130)dB, (31.5Hz~12.5kHz)	$U=0.2\text{dB}$		
				Toneburst Signal:(30~130)dB, (0.25~1000)ms	$U=0.3\text{dB}$		
				F:(20.0~50.0)dB/s	$U=3.2\text{dB/s}$		

No. CNAS L0260

第 4 页 共 5 页

The scope of the accreditation in Chinese remains the definitive version.



Nº	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date				
7	Acoustic calibrator	and S	V.R.of Acoustic calibrator JJG176	S:(1.0~8.0)dB/s (80~130)dB,(20~8000)Hz (20~8000)Hz (0.03~10.0)%	U=0.3dB/s U=0.12 dB U _{rel} =0.2% U=0.40%						
		SPL									
		Frequency									
III、Electromagnetic measuring instrument											
1	Fluxgate Magnetometer	Magnetic Induction	Calibration Specification for Fluxgate Magnetometer JJF1519	(-100 μ T~100 μ T)	U=0.3%R _d +10nT						
		Orthogonality		0~10°	U=0.15°						
Special measuring instruments											
IV、Special measuring instruments for construction and traffic											
1	Pile Dynamic Measuring Instrument	Acceleration	V.R.of Pile Dynamic Measuring InstrumentJJG930-1998 JJG930	(0.5~1000)m/s ²	U _{rel} =1.4%						
V、Other measuring instruments											
1	MEMS Gyroscopes	Angular speed	Calibration Specification for MEMS Gyroscopes JJF 1535	(0.01~500)° /s, (-55~85)°C	U _{rel} =1.1×10 ⁻³		认可证书专用章				
				(>500~1500)° /s	U _{rel} =1.1×10 ⁻³						
		Scale Factor		(1~1×10 ⁶)mV/(° /s)	U _{rel} =1.4×10 ⁻³						



No. CNAS L0260

第 5 页 共 5 页

The scope of the accreditation in Chinese remains the definitive version.